

**ATTACHMENT 7
 AUTOMATED EXTERNAL DEFIBRILLATOR
 AED DAILY/MONTHLY SAFETY INSPECTION RECORD FOR:
 FISCAL YEAR 2017**

AED MANUFACTURER and MODEL: **CARDIAC SCIENCE AED**

AED SERIAL NUMBER: **4025384**

AED LOCATION: **GPLS HAGATNA – 1st FLOOR**

Date:MONTHLY	INSPECTORS INITIALS	CARRYING CASE INTACT	BATTERY CHARGED	ADDTL EQUIPMENT	COMMENTS/NOTES
October 2016	JA	YES	YES	NO	NONE
November 2016	JA	YES	YES	NO	NONE
December 2016	JA	YES	YES	NO	NONE
January 2017	JA	YES	YES	NO	NONE
February 2017	JA	YES	YES	NO	NONE
March 2017	JA	YES	YES	NO	NONE
April 2017	JA	YES	YES	NO	NONE
May 2017	JA	YES	YES	NO	NONE
June 2017	JA	YES	YES	NO	NONE
July 2017					
July 2017					
August 2017					
September 2017					

AED INSPECTIONS ARE CONDUCTED MONTHLY

Additional Notes/Comments:

Safety 1st Systems is the company that services the AED machine with both parts and labor.

Things to remember about the machine: Green light is: OK (in working order). Once the green light goes out & the machine starts beeping we need to contact Safety 1st.

EXPIRATION OF DEFIBRILLATOR PADS: 6/17/2018

Prepared & Inspected by:

June M. Aflague
 Administrative Assistant